

## Sky Xtra Face Mask: Key Performance characteristics

Performance of the Sky Xtra Face Mask against the functional particle filtration and breathability performance requirements of popular standards for Face Masks has been independently determined as follows\*:

Tests performed	FFP2	Comparable N95 performance*
<b>Filtration</b> EN 149:2001+A1:2009, Clause 8.11 & AFNOR-SPEC-S76-001:2020, Reference to EN13274-7: 2019 Modified	PASS	PASS
<b>Breathability</b> EN 149:2001+A1:2009, Clause 8.9 & EN ISO 9237-1995	PASS	PASS

### Testing against FFP2 functional performance requirements

The Sky Xtra Face Mask has been independently tested by NTEK against the functional performance requirements of the FFP2 standard and determined to have the following key characteristics when new:

	Requirement	Result**	
<b>Penetration of Filter Material</b> (EN 149:2001+A1:2009, Clause 8.11)	<i>Maximum penetration of test aerosol:</i> Sodium chloride @ 95 L/m ≤ 6% Paraffin oil @ 95 L/m ≤ 6%	Sodium chloride ≤ 2.07% Paraffin oil ≤ 4.39%	PASS
<b>Breathing Resistance</b> (EN 149:2001+A1:2009, Clause 8.9)	<i>Maximum permitted resistance (mbar):</i> Inhalation @ 30 L/min ≤ 0.7 Inhalation @ 95 L/min ≤ 2.4 Exhalation @ 160 L/min ≤ 3.0	Inhalation @ 30 L/min ≤ 0.4 Inhalation @ 95 L/min ≤ 1.46 Exhalation @ 160 L/min ≤ 1.27	PASS
<b>Total Inward Leakage</b> (EN 149:2001+A1:2009 Clause 8.5)	Total inward leakage ≤ 8%	Total inward leakage < 8%	PASS

\*\*NTEK test reports included as appendix

## Comparable N95 performance level\*

	FFP2 Requirement	Comparable N95 Requirement*	Result*
<b>Filter performance</b>	<i>Maximum penetration of test aerosol:</i> Sodium chloride @ 95 L/m ≤ 6% Paraffin oil @ 95 L/m ≤ 6%	<i>Maximum penetration of test aerosol:</i> Sodium chloride @ 85 L/m ≤ 5%	<b>Sodium chloride ≤ 2.07%</b> <b>Paraffin oil ≤ 4.39%</b>
<b>Breathing Resistance</b>	<i>Maximum permitted resistance (mbar):</i> Inhalation @ 30 L/min ≤ 0.7 Inhalation @ 95 L/min ≤ 2.4 Exhalation @ 160 L/min ≤ 3.0	<i>Maximum permitted resistance (mbar):</i> Inhalation @ 85 L/min ≤ 3.43 Exhalation @ 85 L/min ≤ 2.45	Inhalation @ 30 L/min ≤ 0.4 <b>Inhalation @ 95 L/min ≤ 1.46</b> <b>Exhalation @ 160 L/min ≤ 1.27</b>

Refer to <https://multimedia.3m.com/mws/media/17915000/comparison-ffp2-kn95-n95-filtering-facepiece-respirator-classes-tb.pdf> for a helpful comparison between FFP2, N95 and other international standards.

The test results for the Sky Xtra Face Mask are presented on the following pages.

\* Comparisons between standards are for illustrative purposes only.

Mask has not been FDA cleared or approved.

**Flashbay**  
February 2021

# Test Report

**Applicant:** Flashbay Electronics  
**Address:** Building 2, Jixun Industrial Park, Xinjiao, Dong'ao Village, Shatian Town,  
Huiyang District, Huizhou City, Guangdong Province, P.R.China

**The following sample(s) was/were submitted and identified on behalf of the client as:**



Product name: Face Mask  
Model: Sky Xtra(SKX)  
Manufacturer: Flashbay Electronics  
Address: Building 2, Jixun Industrial Park, Xinjiao, Dong'ao Village, Shatian Town,  
Huiyang District, Huizhou City, Guangdong Province, P.R.China  
Classification: FFP2 NR  
Sample quantity: 30 Pcs  
Sample Received Date: Feb. 04, 2021  
Testing Period: Feb. 04, 2021~ Feb. 22, 2021

**Test Requirement:**

According to the requirement of the client, the test item(s) of the sample is referring to the standard EN 149:2001+A1:2009.

**Test Result(s):** Please refer to the following page(s)

**Test Method:** Please refer to the following page(s)

Compiled by:  Reviewed by:   
Approved by:  Date: 2021-02-23

## Test Result

### Clause 7.9.2 Penetration of Filter Material

(EN 149:2001+A1:2009, Clause 8.11)

Test Requirement			Results
The penetration of the filter of the particle filtering half mask shall meet the requirements of the following table.			Detail refer to Appendix 1
Classification	Maximum penetration of test aerosol(%)		
	Sodium chloride test 95 L/min	Paraffin oil test 95 L/min	
FFP1	20	20	
FFP2	6	6	
FFP3	1	1	

### Appendix 1: Summarization of Test Data

Penetration of filter material

Aerosol	Condition	Sample No.	Penetration (%)	
			Average in 30s after 3 min	Max. during exposure
Sodium chloride test	A.R.	1#	2.07	/
		2#	1.64	/
		3#	1.19	/
Paraffin oil test	A.R.	4#	4.38	/
		5#	3.86	/
		6#	4.39	/

Flow rate of test aerosol: 95.0 L/min

**Clause 7.9.1 Total Inward Leakage**

(EN 149:2001+A1:2009 Clause 8.5)

Test Requirement	Results
<p>For particle filtering half masks fitted in accordance with the manufacturer's information, at least 46 out of the 50 individual exercise results (i.e. 10 subjects x 5 exercises) for total inward leakage shall be not greater than:</p> <p style="padding-left: 40px;">25% for FFP1 11% for FFP2 5% for FFP3</p> <p>and, in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than:</p> <p style="padding-left: 40px;">22% for FFP1 8% for FFP2 2% for FFP3</p>	<p>Detail refer to Appendix 2</p>

**Appendix 2: Summarization of Test Data**

Subject	Sample	Condition	Normal Breathing (%)	Head Side/Side (%)	Head Up/Down (%)	Speak Loudly (%)	Normal Breathing (%)	Mean (%)
Gu	7#	A.R.	7.2	7.3	7.5	7.6	7.3	7.38
Hu	8#	A.R.	6.8	6.9	7.2	7.4	6.9	7.04
Wang	9#	A.R.	6.5	6.6	6.7	6.8	6.6	6.64
Long	10#	A.R.	7.4	7.6	7.7	7.9	7.5	7.62
Gao	11#	A.R.	6.9	7.1	7.2	7.4	7.1	7.14
Huang	15#	A.R.	6.9	7.1	7.2	7.3	7.1	7.12
Zhou	16#	A.R.	5.2	5.4	5.6	5.7	5.3	5.44
Ma	17#	A.R.	7.2	7.3	7.4	7.6	7.4	7.38
Wu	18#	A.R.	7.5	7.7	7.8	7.9	7.6	7.70
Li	19#	A.R.	6.2	6.3	6.4	6.6	6.4	6.38

Facial Dimension:

Subject	Length of Face ( mm )	Width of Face ( mm )	Depth of Face ( mm )	Width of Mouth ( mm )
Gu	114	127	119	52
Hu	128	144	135	53
Wang	112	136	122	50
Long	119	134	128	51
Gao	130	154	144	52
Huang	130	140	125	53
Zhou	100	148	125	55
Ma	120	158	110	50
Wu	110	148	121	44
Li	112	146	112	50

**Clause 7.16 Breathing Resistance**

EN 149:2001+A1:2009, Clause 8.9)

Test Requirement				Results
The breathing resistances apply to valved and valveless filtering half masks and shall meet the requirements as the following table.				Detail refer to Appendix 3
Classification	Maximum permitted resistance (mbar)			
	Inhalation		Exhalation	
	30 L/min	95 L/min	160 L/min	
FFP1	0.6	2.1	3.0	
FFP2	0.7	2.4	3.0	
FFP3	1.0	3.0	3.0	

**Appendix 3: Summarization of Test Data**

Specimen	Condition	Inhalation(mbar)		Exhalation resistance(mbar)				
		At 30 L/min	At 95 L/min	At 160 L/min				
				A	B	C	D	E
12#	A.R.	0.38	1.43	1.25	1.26	1.24	1.25	1.25
13#		0.39	1.45	1.26	1.25	1.26	1.26	1.25
14#		0.40	1.46	1.26	1.25	1.26	1.27	1.26

A: facing directly ahead; B: facing vertically upwards; C: facing vertically downwards; D: lying on the left side; E: lying on the right side

**Remark:**

According to the requirement of the client, only the specimen of "A.R." has been tested.

**Sample photo(s):**



Fig.1



Fig.2

This testing report displaces the original report of No. S21020400101E, and the original one No. S21020400101E was invalid since the date of this testing report released.

\*\*\*\*End of Report\*\*\*\*

The test report is effective only with both signature and specialized stamp, the result(s) shown in this report refer only to the sample(s) tested. Without written approval of NTEK, this report can't be reproduced except in full; The laboratory is not responsible for the authenticity of the sample information provided by the customer; The laboratory is not responsible for any deviation of results due to methods/standards provided by the customer.